






Product specifications

To learn more about our contact lenses visit:
www.cooper-vision.co.uk



Product	Sphere Power (DS) [†]	Cylinder Power (DC)	Axis°	Add Power (D)	Design	Design Technology	Wear Schedule & Replacement Frequency	Base Curve (mm)	Diameter (mm)	Material Technology	Material USAN	Water Content (%)	Oxygen Transmissibility Dk/t [‡]	UV Blocking [†] Class	Visibility Tint	Material Group
 MiSight® 1 day	-10.00 to -6.50 (0.50 steps) -6.00 to -0.25 (0.25 steps)	— — —	— — —	Treatment zones offer +2.00 of myopic defocus	Dual-focus	ActivControl® Technology	Daily wear; one-day replacement	8.7	14.2	PC Technology® (phosphoryl choline)	omafilcon A	60	28	No	Yes	2
 MyDay®	-12.00 to -6.50 (0.50 steps) -6.00 to -0.25 (0.25 steps) +0.25 to +5.00 (0.25 steps) +5.50 to +8.00 (0.50 steps)	— — —	— — —	— — —	Asphere	Aberration Neutralising System™	Daily wear; one-day replacement	8.4	14.2	Aquaform® Technology	stenfilcon A	54	100	Class 2	Yes	5B (SiH)
 MyDay® toric	-10.00 to -6.50 (0.50 steps) -6.00 to Plano (0.25 steps) +0.25 to +6.00 (0.25 steps) +6.50 to +8.00 (0.50 steps)	-0.75 -1.25 -1.75 -2.25	10 to 180 (in 10 steps)	— — —	Toric	Optimised Toric Lens Geometry™	Daily wear; one-day replacement	8.6	14.5	Aquaform® Technology	stenfilcon A	54	80	Class 2	Yes	5B (SiH)
 MyDay® multifocal	-12.00 to -10.50 (0.50 steps) -10.00 to +8.00 (0.25 steps)	— — —	— — —	Low (+0.75 to +1.25) Med (+1.50 to +1.75) High (+2.00 to +2.50)	Multifocal	Binocular Progressive System™	Daily wear; one-day replacement	8.4	14.2	Aquaform® Technology	stenfilcon A	54	100	Class 2	Yes	5B (SiH)
 MyDay® Energys (US only 2023)	-12.00 to -6.50 (0.50 steps) -6.00 to -0.25 (0.25 steps) +0.25 to +5.00 (0.25 steps) +5.50 to +8.00 (0.50 steps)	— — —	— — —	— — —	Asphere	DigitalBoost	Daily wear; one-day replacement	8.4	14.2	Aquaform® Technology	stenfilcon A	54	100	Class 2	Yes	5B (SiH)

[†] Plano lens availability for sphere product can vary by market or customer.

[‡] (@-3.00DS) x 10⁻⁹ [(cm/sec) x (ml O₂)/(ml x mmHg)]





[†] UV-blocking contact lenses help provide protection against transmission of harmful UV radiation to eye but are not substitutes for protective UV-absorbing eyewear, such as UV-absorbing goggles or sunglasses, as they do not completely cover the eye or surrounding area. Continue to use UV-absorbing eyewear as directed by your eye care professional.

Packaging images for illustration purposes only.

Product specifications

To learn more about our contact lenses visit:
www.cooper-vision.co.uk



Product	Sphere Power (DS) [*]	Cylinder Power (DC)	Axis ^o	Add Power (D)	Design	Design Technology	Wear Schedule & Replacement Frequency	Base Curve (mm)	Diameter (mm)	Material Technology	Material USAN	Water Content (%)	Oxygen Transmissibility Dk/t [†]	UV Blocking [‡] Class	Visibility Tint	Material Group
 clariti [®] 1 day sphere	-10.00 to -6.50 (0.50 steps) -6.00 to -0.50 (0.25 steps) +0.50 to +6.00 (0.25 steps) +6.50 to +8.00 (0.50 steps)	— — —	— — —	— — —	Asphere	— — —	Daily wear; one-day replacement	8.6	14.1	WetLoc [®] Technology	somofilcon A	56	86	Class 2	No	5B (SIH)
 clariti [®] 1 day toric	-9.00 to -6.50 (0.50 steps)	-0.75 -1.25 -1.75	10, 20, 60, 70, 80, 90, 100, 110, 120, 160, 170, 180	— — —	Toric	Smooth-gradient ballast toric design	Daily wear; one-day replacement	8.6	14.3	WetLoc [®] Technology	somofilcon A	56	57	Class 2	No	5B (SIH)
		-2.25	10, 20, 90, 160, 170, 180													
	-6.00DS to Plano (0.25 steps)	-2.25	10, 20, 70, 80, 90, 100, 110, 160, 170, 180													
		-0.75 -1.25 -1.75	10 to 180 (10 steps)													
+0.25 to +4.00 (0.25 steps)	-0.75 -1.25 -1.75	10, 20, 70, 80, 90, 100, 110, 160, 170, 180														
 clariti [®] 1 day multifocal	-6.00 to +5.00 (0.25 steps)	— — —	— — —	Low High	Multifocal	— — —	Daily wear; one-day replacement	8.6	14.1	WetLoc [®] Technology	somofilcon A	56	86	Class 2	No	5B (SIH)
 Live [®]	-10.00 to -6.50 (0.50 steps)	— — —	— — —	— — —	Asphere	— — —	Daily wear; one-day replacement	8.6	14.0	AquaGen [®] Technology	somofilcon A	56	86	Class 2	No	5B (SIH)
	-6.00 to -0.50 (0.25 steps)															
	+0.50 to +6.00 (0.25 steps)															
	+6.50 to +8.00 (0.50 steps)															

* Plano lens availability for sphere product can vary by market or customer.

[†] (@-3.00DS) x 10⁻⁹ [(cm/sec) x (ml O₂)/(ml x mmHg)]





[‡] UV-blocking contact lenses help provide protection against transmission of harmful UV radiation to eye but are not substitutes for protective UV-absorbing eyewear, such as UV-absorbing goggles or sunglasses, as they do not completely cover the eye or surrounding area. Continue to use UV-absorbing eyewear as directed by your eye care professional.

Packaging images for illustration purposes only.

Product specifications

To learn more about our contact lenses visit:
www.cooper-vision.co.uk



Product	Sphere Power (DS)*	Cylinder Power (DC)	Axis°	Add Power (D)	Design	Design Technology	Wear Schedule & Replacement Frequency	Base Curve (mm)	Diameter (mm)	Material Technology	Material USAN	Water Content (%)	Oxygen Transmissibility Dk/t†	UV Blocking‡ Class	Visibility Tint	Material Group
 Proclear® 1 day	-12.00 to -6.50 (0.50 steps) -6.00 to -0.25 (0.25 steps) +0.25 to +5.00 (0.25 steps) +5.50 to +8.00 (0.50 steps)	— — —	— — —	— — —	Asphere	Aberration Neutralising System™	Daily wear; one-day replacement	8.7	14.2	PC Technology® (phosphoryl choline)	omafilcon A	60	28	No	Yes	2
 Proclear® 1 day multifocal	-10.00 to -6.50 (0.50 steps) -6.00 to +6.00 (0.25 steps)	— — —	— — —	Single power profile +1.50D (Could fit adds up to +2.50)	Multifocal	— — —	Daily wear; one-day replacement	8.7	14.2	PC Technology® (phosphoryl choline)	omafilcon A	60	28	No	Yes	2
 Biomedics® 1 day Extra	-10.00 to -6.50 (0.50 steps) -6.00 to -0.25 (0.25 steps) +0.25 to +5.00 (0.25 steps) +5.50 to +6.00 (0.50 steps)	— — —	— — —	— — —	Sphere	— — —	Daily wear; one-day replacement	8.6 8.8	14.2	— — —	ocufilcon D	55	26	No	Yes	4
 Biomedics® 1 day Extra toric	-10.00 to -6.50 (0.50 steps) -6.00 to Plano (0.25 steps)	-0.75 -1.25 -1.75	20°, 90°, 160°, 180° (Plano to -7.00DS) and 90°, 180° (-7.50 to -10.00DS)	— — —	Toric	Optimised Toric Lens Geometry®	Daily wear; one-day replacement	8.7	14.5	— — —	ocufilcon D	55	18	No	Yes	4

* Plano lens availability for sphere product can vary by market or customer.

† (@-3.00DS) x 10⁻⁹ [(cm/sec) x (ml O₂)/(ml x mmHg)]





‡ UV-blocking contact lenses help provide protection against transmission of harmful UV radiation to eye but are not substitutes for protective UV-absorbing eyewear, such as UV-absorbing goggles or sunglasses, as they do not completely cover the eye or surrounding area. Continue to use UV-absorbing eyewear as directed by your eye care professional.

Packaging images for illustration purposes only.

Product specifications

To learn more about our contact lenses visit:
www.coopervision.co.uk



Product	Sphere Power (DS)*	Cylinder Power (DC)	Axis°	Add Power (D)	Design	Design Technology	Wear Schedule & Replacement Frequency	Base Curve (mm)	Diameter (mm)	Material Technology	Material USAN	Water Content (%)	Oxygen Transmissibility Dk/t†	UV Blocking‡ Class	Visibility Tint	Material Group
 Biofinity®	-12.00 to -6.50 (0.50 steps) -6.00 to -0.25 (0.25 steps) +0.25 to +6.00 (0.25 steps) +6.50 to +8.00 (0.50 steps) NO PLANO	— — —	— — —	— — —	Asphere	Aberration Neutralising System™	Daily wear; 30 days replacement Extended wear 7 days/6 nights; 30 day replacement	8.6	14.0	Aquaform® Technology	comfilcon A	48	171	No	Yes	5C (SiH)
 Biofinity® XR	-20.00 to -12.50 (0.50 steps) +8.50 to +15.00 (0.50 steps)	— — —	— — —	— — —	Asphere	Aberration Neutralising System™	As for Biofinity®	8.6	14.0	Aquaform® Technology	comfilcon A	48	171	No	Yes	5C (SiH)
 Biofinity Energys®	-12.00 to -6.50 (0.50 steps) -6.00 to -0.25 (0.25 steps) +0.25 to +6.00 (0.25 steps) +6.50 to +8.00 (0.50 steps) NO PLANO	— — —	— — —	— — —	Asphere	Digital Zone Optics®	As for Biofinity®	8.6	14.0	Aquaform® Technology	comfilcon A	48	171	No	Yes	5C (SiH)
 Biofinity® toric	-10.00 to -6.50 (0.50 steps) -6.00 to plano (0.25 steps) +0.25 to +6.00 (0.25 steps) +6.50 to +8.00 (0.50 steps)	-0.75 -1.25 -1.75 -2.25	10 to 180 (10 steps)	— — —	Toric	Optimised Toric Lens Geometry™	As for Biofinity®	8.7	14.5	Aquaform® Technology	comfilcon A	48	116	No	Yes	5C (SiH)

* Plano lens availability for sphere product can vary by market or customer.

† (@-3.00DS) x 10⁻⁹ [(cm/sec) x (ml O₂)/(ml x mmHg)]




‡ UV-blocking contact lenses help provide protection against transmission of harmful UV radiation to eye but are not substitutes for protective UV-absorbing eyewear, such as UV-absorbing goggles or sunglasses, as they do not completely cover the eye or surrounding area. Continue to use UV-absorbing eyewear as directed by your eye care professional.

Packaging images for illustration purposes only.

Product specifications

To learn more about our contact lenses visit:
www.coopervision.co.uk



Product	Sphere Power (DS) [†]	Cylinder Power (DC)	Axis°	Add Power (D)	Design	Design Technology	Wear Schedule & Replacement Frequency	Base Curve (mm)	Diameter (mm)	Material Technology	Material USAN	Water Content (%)	Oxygen Transmissibility Dk/t [‡]	UV Blocking [‡] Class	Visibility Tint	Material Group
 Biofinity® XR Toric	-20.00 to -10.50 (0.50 steps) +8.50 to +20.00 (0.50 steps)	-0.75 -1.25 -1.75 -2.25	5 to 180 (5 steps)	— —	Toric	Optimised Toric Lens Geometry™	As for Biofinity sphere	8.7	14.5	Aquaform® Technology	comfilcon A	48	116	No	Yes	5C (SiH)
	-20.00 to -6.50 (0.50 steps) -6.00 to plano (0.25 steps) +0.25 to +6.00 (0.25 steps) +6.50 to +20.00 (0.50 steps)	-2.75 -3.25 -3.75 -4.25 -4.75 -5.25 -5.75														
 Biofinity® multifocal	-10.00 to -6.50 (0.50 steps) -6.00 to plano (0.25 steps) +0.25 to +6.00 (0.25 steps)	— —	— —	+1.00 +1.50 +2.00 +2.50	Multifocal D lens N Lens	Balanced Progressive® Technology	As for Biofinity sphere	8.6	14.0	Aquaform® Technology	comfilcon A	48	142 (-3.00, N lens, +1.00 Add)	No	Yes	5C (SiH)
 Biofinity® toric multifocal	-10.00 to -6.50 (0.50 steps) -6.00 to plano (0.25 steps) +0.25 to +6.00 (0.25 steps) +6.50D to +10.00 (0.50 steps)	-0.75 to -5.75 (0.50 steps)	5 to 180 (5 steps)	+1.00 +1.50 +2.00 +2.50	Toric Multifocal D Lens N Lens	Optimised Toric Lens Geometry™ and Balanced Progressive® Technology	As for Biofinity sphere	8.7	14.5	Aquaform® Technology	comfilcon A	48	116	No	Yes	5C (SiH)





[†] (@-3.00DS) x 10⁻⁹ [(cm/sec) x (ml O₂)/(ml x mmHg)]

[‡] UV-blocking contact lenses help provide protection against transmission of harmful UV radiation to eye but are not substitutes for protective UV-absorbing eyewear, such as UV-absorbing goggles or sunglasses, as they do not completely cover the eye or surrounding area. Continue to use UV-absorbing eyewear as directed by your eye care professional. Packaging images for illustration purposes only.

Product specifications

To learn more about our contact lenses visit:
www.coopervision.co.uk



Product	Sphere Power (DS)*	Cylinder Power (DC)	Axis°	Add Power (D)	Design	Design Technology	Wear Schedule & Replacement Frequency	Base Curve (mm)	Diameter (mm)	Material Technology	Material USAN	Water Content (%)	Oxygen Transmissibility Dk/t†	UV Blocking‡ Class	Visibility Tint	Material Group
 Avaira Vitality®	-12.00 to -6.50 (0.50 steps) -6.00 to -0.25 (0.25 steps) +0.25 to +6.00 (0.25 steps) +6.50 to +8.00 (0.50 steps) NO PLANO	— — —	— — —	— — —	Asphere	Aberration Neutralising System™	Daily wear; 30 days replacement	8.4	14.2	— — —	fanfilcon A	55	112	Class 1	Yes	5B (SiH)
 Avaira Vitality® toric	-10.00 to -6.50D (0.50D steps) -6.00 to Plano (0.25D steps) +0.25 to +6.00D (0.25D steps) +6.50 to +8.00D (0.50D steps)	-0.75 -1.25 -1.75 -2.25	10 to 180 (10 steps)	— — —	Toric	Optimised Toric Lens Geometry™	Daily wear; 30 days replacement	8.5	14.5	— — —	fanfilcon A	55	90	Class 1	Yes	5B (SiH)
 Proclear®	-20.00 to -6.50 (0.50 steps) -6.00 to -0.25 (0.25 steps) +0.50 to +6.00 (0.25 steps) +6.50 to +20.00 (0.50 steps) NO PLANO	— — —	— — —	— — —	Sphere	— — —	Daily wear; 30 days replacement	8.6	14.2	PC Technology® (phosphoryl choline)	omafilcon B	62	36	No	Yes	2
 Proclear® toric	-8.00 to -7.00 (0.50 steps) -6.50 to Plano (0.25 steps) +0.25 to +6.00 (0.25 steps)	-0.75 -1.25 -1.75 -2.25	10 to 180 (10 steps)	— — —	Toric	— — —	Daily wear; 30 days replacement	8.4 (MTO) 8.8 (MTO or moulded)	14.4	PC Technology® (phosphoryl choline)	omafilcon B	62	16 (MTO) 19.4 (MTO) 23.7 (Moulded)	No No	Yes Yes	2 2

* Plano lens availability for sphere product can vary by market or customer.

† (@-3.00DS) x 10⁻⁹ [(cm/sec) x (ml O₂)/(ml x mmHg)]





‡ UV-blocking contact lenses help provide protection against transmission of harmful UV radiation to eye but are not substitutes for protective UV-absorbing eyewear, such as UV-absorbing goggles or sunglasses, as they do not completely cover the eye or surrounding area. Continue to use UV-absorbing eyewear as directed by your eye care professional.

Packaging images for illustration purposes only.

Product specifications

To learn more about our contact lenses visit:
www.coopervision.co.uk



Product	Sphere Power (DS)*	Cylinder Power (DC)	Axis°	Add Power (D)	Design	Design Technology	Wear Schedule & Replacement Frequency	Base Curve (mm)	Diameter (mm)	Material Technology	Material USAN	Water Content (%)	Oxygen Transmissibility Dk/t†	UV Blocking‡ Class	Visibility Tint	Material Group
 Proclear® XR toric	-10.00 to -7.00 (0.50 steps) -6.50 to Plano (0.25 steps) +0.25 to +6.50 (0.25 steps) +7.00 to +10.00 (0.50 steps)	-0.75 to -5.75D (0.50D steps)	5 to 180 (5 steps)	---	Toric	---	Daily wear; 30 days replacement	8.8	14.4	PC Technology® (phosphoryl choline)	omafilcon B	62	13.9	No	Yes	2
 Proclear® multifocal	-8.00 to -7.00 (0.50 steps) -6.50 to Plano (0.25 steps) +0.25 to +6.00 (0.25 steps)	---	---	+1.00 +1.50 +2.00 +2.50	Multifocal D lens N lens	---	Daily wear; 30 days replacement	8.7	14.4	PC Technology® (phosphoryl choline)	omafilcon B	62	14 (MTO)	No	Yes	2
 Proclear® XR multifocal	-20.00 to -7.00 (0.50 steps) -6.50 to Plano (0.25 steps) +0.25 to +6.50 (0.25 steps) +7.00 to +20.00D (0.50 steps)	---	---	+1.00 to +4.00 (0.50 steps)	Multifocal D lens N lens	---	Daily wear; 30 days replacement	8.7	14.4	PC Technology® (phosphoryl choline)	omafilcon B	62	From: 14.0 (-3.00D/+4.00D, N lens) To: 16.7 (-3.00D/+2.50D, D lens)	No	Yes	2
 Proclear® multifocal toric	-20.00 to -7.00 (0.50 steps) -6.50 to Plano (0.25 steps) +0.25 to +6.50 (0.25 steps) +7.00 to +20.00D (0.50 steps)	-0.75 to -5.75 (0.50 steps)	5 to 180 (5 steps)	+1.00 to +4.00 (0.50 steps)	Toric Multifocal D lens N lens	---	Daily wear; 30 days replacement	8.4 8.8	14.4	PC Technology® (phosphoryl choline)	omafilcon B	62	From: 14.5 (-3.00D/+4.00D, N lens) To: 17.7 (-3.00D/+4.00D, D lens)	No	Yes	2

* Plano lens availability for sphere product can vary by market or customer.

† (@-3.00DS) x 10⁻⁹ [(cm/sec) x (ml O₂)/(ml x mmHg)]





‡ UV-blocking contact lenses help provide protection against transmission of harmful UV radiation to eye but are not substitutes for protective UV-absorbing eyewear, such as UV-absorbing goggles or sunglasses, as they do not completely cover the eye or surrounding area. Continue to use UV-absorbing eyewear as directed by your eye care professional.

Packaging images for illustration purposes only.

Product specifications

To learn more about our contact lenses visit:
www.cooper-vision.co.uk



Product	Sphere Power (DS) [*]	Cylinder Power (DC)	Axis [°]	Add Power (D)	Design	Design Technology	Wear Schedule & Replacement Frequency	Base Curve (mm)	Diameter (mm)	Material Technology	Material USAN	Water Content (%)	Oxygen Transmissibility Dk/t [†]	UV Blocking [‡] Class	Visibility Tint	Material Group
 Biomedics® 55 Evolution	-10.00 to -6.50 (0.50 steps)	---	---	---	Asphere	Aberration Neutralising System™	Daily wear; 30 days replacement	8.6, 8.9	14.2	---	ocufilcon D	55	26	Class 2	Yes	4
	-6.00 to -0.25 (0.25 steps)							8.8								
 Biomedics® toric	+0.25 to +5.00 (0.25 steps)	---	---	---	Asphere	Aberration Neutralising System™	Daily wear; 30 days replacement	8.8	14.2	---	ocufilcon D	55	26	Class 2	Yes	4
	+5.50 to +8.00 (0.50 steps)															
	NO PLANO															
 Biomedics® toric	-9.00 to -6.50 (0.50 steps)	-0.75 -1.25 -1.75 -2.25	10 to 180 (10 steps)	---	Toric	---	Daily wear; 30 days replacement	8.7	14.5	---	ocufilcon D	55	18	Class 2	Yes	4
	-6.00 to Plano (0.25 steps)															
	+0.25 to +5.00 (0.25 steps)															
 Biomedics® toric	+5.00 to +6.00 (0.50 steps)	---	---	---	Toric	---	Daily wear; 30 days replacement	8.7	14.5	---	ocufilcon D	55	18	Class 2	Yes	4

^{*} Plano lens availability for sphere product can vary by market or customer.

[†] (@-3.00DS) x 10⁻⁹ [(cm/sec) x (ml O₂)/(ml x mmHg)]

[‡] UV-blocking contact lenses help provide protection against transmission of harmful UV radiation to eye but are not substitutes for protective UV-absorbing eyewear, such as UV-absorbing goggles or sunglasses, as they do not completely cover the eye or surrounding area. Continue to use UV-absorbing eyewear as directed by your eye care professional.

Packaging images for illustration purposes only.

Aberration Neutralising System™, ActivControl Technology®, Aquaform®, AquaGen®, Avaira Vitality®, Biofinity®, Biofinity Energys®, Biomedics®, Binocular Progressive Technology™, clariti®, CooperVision®, Live®, MiSight®, MyDay®, Optimised Toric Lens Geometry®, PC Technology®, Proclear® and WetLoc® are trademarks and registered trademarks of CooperVision UK part of The Cooper Companies.